

TYPE CERTIFICATE



Czech

Registration Number 00.270.885/09/07/02/0

issued for the producer:

INGOS s.r.o.
K Nouzovu 2090
CZ - 143 00 Praha 4
Ident.-No.: 645 80 903

for the product:

Name: **Automatic Amino Acid Analyser**
 Type designation: **AAA 400**
 Manufacturer: **K Nouzovu 2090, 143 16 Praha 4**

at which the certification was carried out according to the certification schedule instruction ISO/IEC 67 - system 3 v conforming with the certification system of the TÜV SÜD Czech which results are given in the Evaluation Report No 90/90/09/BT/IZ/S from Januar 27, 2009.

The mentioned product type fulfils the applicable requirements of following regulations/ normative documents, which have underlain for its evaluation:

Decree of the Government no 17/2003 Coll. (Directive 2006/95/EC)
 ČSN EN 61010-1:2003

Decree of the Government no 616/2006 Coll. (Directive 2004/108/EC)

ČSN EN 61000-6-1:2007, ČSN EN 61000-6-3 ed.2:2007,
 ČSN EN 55011 ed.2:2007, ČSN EN 61000-3-2 ed.3:2006,
 ČSN EN 61000-4-2:1997+Z1:2001, ČSN EN 61000-4-3 ed. 3:2006,
 ČSN EN 61000-4-4 ed.2:2005, ČSN EN 61000-4-5 ed.2:2007,

Decree of the Government no 24/2003 Coll. (Directive 98/37/EC)
 ČSN EN ISO 12100-1:2004, ČSN EN ISO 12100-2:2004, ČSN EN ISO
 13857:2008, ČSN EN 982:1997+ A1:2008, ČSN EN ISO 13732:2007

This certificate is valid till: Februar 05, 2014

Details and validity conditions are given in the appendix to this certificate, which is its integral part and contains 1 page.

Prague, Februar 05, 2009




 Manager of the certification
 body

Product Certification Body