

CERTIFICATE ◆ CERTIFICADO ◆ CERTIFIKAT ◆ 認証証書 ◆ CERTIFICATE ◆ CERTIFIKAT

# TYPE CERTIFICATE



Registration Number **044/06/07/02/0**

issued for the producer:

**INGOS s.r.o.**  
**K Nouzovu 2090**  
**CZ - 143 16 Praha 4**  
**Ident.-no.: 645 80 903**

for the product:

Identification: **High Pressure Pump**  
Type designation: LCP 5020, LCP 5021, LCP 5080, LCP 5081  
Point of production: K Nouzovu 2090, CZ - 143 16 Praha 4

at which the evaluation, certification and audit of the system quality were carried out intent on the elements ensuring the next product conformity in accordance with the certification system of the TÜV C according to ČSN EN 45011:1998, whose results are given in the Report No 5/90/03/BT/IZ/S from 20.1.2006.

The mentioned product type fulfils the applicable requirements of following regulations/ normative documents, which have underlain for its evaluation:  
**The decree of the government 17/2003 Coll. as amended (Directive 73/23/EEC), ČSN EN 61010-1:2003, The decree of the government 18/2003 Coll. as amended (89/336/EEC), ČSN EN 61000-6-1:2002, ČSN EN 61000-6-3:2002+ A11:2005, ČSN EN 55011:1999+ A2:2003, ČSN EN 61000-3-2 ed.2 :2001+A2:2005, ČSN EN 61000-4-2:1997+Z1:2001, ČSN EN 61000-4-3 ed. 2:2003+ A1, ČSN EN 61000-4-4 ed.2:2005, ČSN EN 61000-4-5:1997+ Z1:2001 The decree of the government 24/2003 Coll. as amended (98/37/EC), ČSN EN ISO 12 100-1:2004, ČSN EN ISO 12 100-2:2004, ČSN EN 294:1993, ČSN EN 982:1997**

This certificate is valid till: **26.1.2009**

Details and validity conditions are given in the appendix to this certificate, which is its integral part and contains 1 page.

Prague, on 26.1.2006



  
**Ing. Ivo Dršťák**  
Manager of the certification body

TÜV CZ s.r.o., Novodvorská 994, 142 21 Praha 4 - ČR IČ: 63987121  
Municipal court in Prague, section C, enclosure No. 38432, entry date: 20.7.1995  
Product Certification Body